

Hach Method 10360
Luminescence Measurement of Dissolved Oxygen in Water and Wastewater and
for Use in the Determination of BOD₅ and cBOD₅
Revision 1.2
September 2011

1.0 Scope and Application

- 1.1 This method is for the measurement of dissolved oxygen (DO) in surface and ground water, and municipal and industrial wastewater.
- 1.2 The method may be used as a replacement for the modified Winkler and membrane electrode procedures for the measurement of DO in wastewater treatment processes such as aeration and biological nutrient basins, effluent outfalls, receiving water, and in Biochemical Oxygen Demand (BOD) determinations where it is desired to perform nondestructive DO measurements.
- 1.3 The method is for use in the United States Environmental Protection Agency's (EPA's) survey and monitoring programs for the measurement of DO and for the determination of BOD and cBOD under the Clean Water Act.
- 1.4 This method is capable of measuring DO in the range of 0.20 to 20 mg/L.
- 1.5 Calibration is by single-point water-saturated air (100% saturation).

2.0 Summary of Method

- 2.1 This luminescence-based sensor procedure measures the light emission characteristics from a luminescence-based reaction that takes place at the sensor-water interface. A light emitting diode (LED) provides incident light required to excite the luminophore substrate. In the presence of dissolved oxygen the reaction is suppressed. The resulting dynamic lifetime of the excited luminophore is evaluated and equated to DO concentration.

3.0 Interferences

- 3.1 There are no known interferences at normal wastewater concentrations that interfere with DO detection and quantification with this method.

4.0 Safety

- 4.1 This method does not address all safety issues associated with its use. The laboratory is responsible for maintaining a safe work environment and a current awareness file of OSHA regulations regarding the safe handling of any chemicals specified in this method. A reference file of material safety data sheets (MSDSs) should be available to all personnel involved in these analyses. Additional information on laboratory safety can be found in References 15.5-15.6.

5.0 Equipment for the Measurement of Dissolved Oxygen

- 5.1 BOD bottle 300-mL with stoppers and plastic caps (Hach # 62016 and 241906)
- 5.2 Magnetic Stirring plate (optional)
- 5.3 Magnetic stirring device (optional)
- 5.4 BOD Bottle, glass stopper, 300-mL (Hach Catalog Number 62106)
- 5.5 Pipette, serological, 1-mL (Hach Catalog Number 919002)
- 5.6 Pipette, serological, 5-mL (Hach Catalog Number 53237)

- 5.7 Pipette, serological, 10-mL (Hach Catalog Number (53238)
- 5.8 Pipette Filler (Hach Catalog Number 1218900)
- 5.9 Meter and LBOD Probe (Hach Catalog Number HQDBOD01, for DO measurement in BOD bottles
- 5.10 Meter and LDO Probe (Hach Catalog Number 8506300 for DO measurement in open containers and water bodies
- 5.12 Temperature controlled environment for BOD bottle incubation, 20 ± 1 °C.

6.0 REAGENTS

- 6.1 Buffer Solution, APHA, for BOD, pH 7.2, Phosphate type, (Hach Catalog Number 43149)
- 6.2 Calcium Chloride Solution, APHA, for BOD (Hach Catalog Number 42849)
- 6.3 Ferric Chloride Solution, APHA, for BOD (Hach Catalog Number (42953)
- 6.4 Glucose-glutamic Acid, Standard Solution, Voluette™ Ampoule, 300-mg/L (150 mg/L glucose and 150 mg/L glutamic Acid), 10 mL (Hach Catalog Number 1486510) or, ezGGA Ampoules, 450 mg/L (225 mg/L Glucose and 225 mg/L Glutamic Acid, (Hach Catalog Number 25144-20)
- 6.5 Magnesium Sulfate Solution, APHA, (Hach Catalog Number (43094)
- 6.6 Nitrification Inhibitor (Hach Catalog Number 2845425)
- 6.7 Potassium Iodide Solution (Hach Catalog Number 1228949)
- 6.8 Sodium Thiosulfate Solution – 0.025 N (Hach Catalog Number 35253.
- 6.9 Sodium Hydroxide Solution, 1 N (Hach Catalog Number 104532)
- 6.10 Sodium Hydroxide Pellets, ACS (Hach Catalog Number 18734)
- 6.11 Starch Indicator (Hach Catalog Number 35253)
- 6.12 Sulfuric Acid Solution, 0.020 N (Hach Catalog Number 104532)
- 6.13 Sulfuric Acid Solution, 1.000 N (Hach Catalog Number 127053)

7.0 Standards

7.1 *Initial LDO/LBOD Probe Calibration*

- 7.1.1 Add approximately 1 inch (2.54 cm) of reagent water to a clean BOD bottle and stopper.
- 7.1.2 Shake vigorously for ~ 10 seconds.
- 7.1.3 Allow for the BOD bottle and its contents to equilibrate to room temperature.
- 7.1.4 The stopper may now be removed from the BOD bottle and the probe inserted for calibration purposes.

- 7.1.5 The luminescence technology for measuring dissolved oxygen is a superior technique from that of Winkler titration and membrane potentiometric measurement and has no interferences associated with oxygen detection process (EPA Validation study, 2004). Therefore, do not adjust the luminescence measurement to that of Winkler or membrane readings.

Note: Section 7.1 is a suggested procedure for the preparation of. Other procedures for the preparation of water-saturated air water-saturated air may be equally effective.

7.2 Calibration Verification, Initial Precision and Recovery, and On-going Precision and Recovery

- 7.2.1 Add approximately 1500 mL of organic-free water or BOD dilution water to a 2-L beaker or PET bottle.
- 7.2.2 Allow the water to equilibrate to room temperature ($\pm 2^\circ \text{C}$).
- 7.2.3 With a steady stream of filtered air ($\approx 10 - 40 \text{ mL per minute}$), aerate the water for a minimum of 30 minutes. Alternatively, vigorously shake the reagent water or BOD dilution water for several minutes.
- 7.2.4 At the completion of aeration, let water re-equilibrate to room temperature ($\pm 2^\circ \text{C}$) for 30 minutes and note the barometric pressure of the laboratory during preparation.
- 7.2.5 Transfer the aerated water to a BOD bottle until overflowing and stopper.
- 7.2.6 Calculate the theoretical dissolved oxygen concentration using a dissolved oxygen table such as Hitchman (1978).

Note: Section 7.2 is a suggested procedure for the preparation of air-saturated water. Other procedures for the preparation of air saturated-water may be equally effective.

8.0 Sample Collection Preservation and Storage

- 8.1 See Title 40 of the Code of Federal Regulations Part 136.3, Table II (Section 15.3) for information regarding required sample collection containers, preservation techniques and holding times for collection of water for measurement of DO and for the determination of BOD and cBOD.

9.0 Quality Control

- 9.1 It is recommended that each laboratory that uses this method be required to operate a formal quality assurance program (Reference 15.1). The minimum requirements of this program consist of an initial demonstration of laboratory capability and ongoing analyses of laboratory prepared water standards as a test of continued performance to assess accuracy and precision. Laboratory performance is compared to established performance criteria to determine if the results of analyses meet the performance characteristics of the method.
- 9.1.1 The analyst shall make an initial demonstration of the ability to generate acceptable accuracy and precision with this method. This ability is established as described in Section 9.2.
- 9.1.2 The laboratory shall, on an ongoing basis, demonstrate through calibration verification and analysis of the ongoing precision and recovery sample that the analysis system is in control. These procedures are described in Sections 9.3 and 9.4, respectively.
- 9.1.3 Accompanying QC for the determination of DO is required per analytical batch. An analytical batch is a set of samples processed during a contiguous 8-hour period. Each analytical

batch should be accompanied by a calibration verification and ongoing precision and recovery sample, resulting in a minimum of three analyses (1 CV, 1 sample, and 1 OPR).

9.2 Initial Demonstration of Laboratory Capability

9.2.1 Initial precision and recovery (IPR) - To establish the ability to generate acceptable precision and accuracy for the measurement of DO in water, the analyst shall perform the following operations:

9.2.1.1 Prepare and measure four samples of the IPR standard (Section 7.2) according to the procedure beginning in Section 10.

9.2.1.2 Using the results of the set of four analyses, compute the average percent recovery (X) and the standard deviation of the percent recovery (s) for DO. Use the following equation for calculation of the standard deviation of the percent recovery:

$$s = \sqrt{\frac{\sum x^2 - \frac{(\sum x)^2}{n}}{n-1}}$$

where:

n = Number of samples

x = Concentration in each sample

9.2.1.3 Compare s and X with the corresponding limits for initial precision and recovery in Tables 5 and 6. If s and X meet the acceptance criteria, system performance is acceptable and analysis of samples may begin. If, however, s exceeds the precision limit or X falls outside the range for recovery, system performance is unacceptable. In this event correct the problem, and repeat the test.

9.3 Calibration Verification

9.3.1 Upon air calibration, prepare a calibration verification standard (Section 7.2) with each analytical batch. Analyze according to the procedure beginning in Section 10 and compare the recovery results to those in Tables 5 and 6.

9.4 Ongoing Calibration and Precision and Recovery

9.4.1 To demonstrate that the analysis system is in control, and acceptable precision and accuracy is being maintained with each analytical batch, the analyst shall perform the following operations

9.4.2 Prepare a precision and recovery standard (Section 7.2) with each analytical batch according to the procedure beginning in Section 10.

9.4.3 Initially, at the end of each analytical batch of samples, analyze a precision and recovery standard and compare the concentration recovery with the limits for ongoing precision and recovery in Tables 5 and 6. If the recovery is in the range specified, measurement process is in control and analysis of samples may proceed. If, however, the recovery is not in the specified range, the analytical process is not in control. In this event, correct the problem,

recalibrate and verify the calibration and reanalyze analytical batch, repeating the ongoing precision and recovery test.

- 9.4.4 The laboratory should add results that pass the specification in Tables 5 and 6 to IPR and previous OPR data and update QC charts to form a graphic representation of continued laboratory performance. The laboratory should also develop a statement of laboratory data quality for each analyte by calculating the average percent recovery (R) and the standard deviation of the percent recovery (sr). Express the accuracy as a recovery interval from $R - 2sr$ to $R + 2sr$. For example, if $R = 95\%$ and $sr = 5\%$, the accuracy is 85% to 105%.
- 9.5 Depending upon specific program requirements, field replicates may be required to assess the precision and accuracy of the sampling and sample transporting techniques.
- 9.6 Glucose-Glutamic Acid Seed Strength Check
- 9.6.1 Many factors can influence the BOD analysis (toxicity from sample matrix, contaminated dilution water, poor quality seed, etc. In order to insure sufficient seeding in the BOD test, a glucose-glutamic acid check is performed in parallel with BOD₅ test. Well prepared dilution water and an active seed will produce a BOD₅ of 198 ± 30 mg/L BOD.

9.6.1.1 Prepare in triplicate a 300-mL BOD bottle with 3.0 mL of the 300 mg/L Standard Solution of GGA (Section 6.4) or 1 ampoule (2.0 mL) of ezGGA (Section 6.4).

9.6.1.2 When using ezGGA ampoules, place the ezGGA ampoule in the ampoule breaker (provided with ezGGA ampoules) and rinse the assembly with reagent water. Hold the ampoule and breaker over the rim of the BOD bottle, break and allow the ampoule to fall into the BOD bottle. Leave the ampoule in the BOD bottle during the incubation period and reading of DO.

Standard Solution

$$\frac{3.0 \text{ mL} \times 0.300 \text{ mg/mL GGA}}{300 \text{ mL final volume}} \times 1000 \text{ mL/L} = 3.0 \text{ mg/L GGA per bottle}$$

ezGGA Ampoule

$$\frac{1 \text{ ampoule (2.0 mL)} \times 0.450 \text{ mg/mL GGA}}{300 \text{ mL final volume}} \times 1000 \text{ mL/L} = 3.0 \text{ mg/L GGA per bottle}$$

9.6.1.2 Add seed at three different volumes (typically 4 mL, 6 mL, and 8 mL,) to the GGA bottles. Other volumes may be required, depending on the strength of the seed being used.

9.6.1.3 Bring to volume with dilution water and analyze as described in Sections 12.9 and 12.10.

Note: GGA BOD₅ recovery results outside of 198 ± 30 mg/L should be investigated as to causation. If toxicity of dilution water has been ruled out as a probable cause for low recovery, it is likely that the seed is of low activity or poor quality. Either increase the seed amount or use a seed of higher quality. High GGA recoveries are generally due to incorrect amount of GGA Standard Solution.

10.0 Calibration and Standardization

- 10.1 Because of the possible diversity of future LDO instrument hardware and, no detailed operating conditions are provided. The analyst is advised to follow the recommended operating conditions provided by the manufacturer. It is the responsibility of the analyst to verify that the instrument

configuration and operating conditions satisfy the analytical requirements of this method and to maintain quality control data verifying instrument performance and analytical results.

10.2 Water-saturated air (Section 7.1) is used for instrument calibration.

10.3 Calibration verification (Section 7.2) is performed with air-saturated water prior to any DO sample measurements to the method specifications.

11.0 Procedure for Measuring DO in Grab Samples, Outfalls, and Open Water Bodies

11.1 Instrument Setup

11.1.1 Follow the instrument manufacturer's instructions for instrument setup (Hach Document DOC022.53.80021 for Hach LDO IntelliCal™ Rugged and Standard Probes, Hach Catalog Number 5790018 for Hach LDO process probes.

Note: Manufacturer's instructions are only for instrument set and use. These instructions do not preclude the calibration and performance requirements in this method.

11.2 Measurement of DO

11.2.1 Insert the LDO probe into the water sample to be measured and stir gently probe with or add a stir bar. Do not put the probe on the bottom or sides of the container. Stir the sample at a moderate rate or put the probe in flowing conditions. Read sample. The display will show "Stabilizing" and a progress bar as the probe stabilizes in the sample. The display will show the lock icon when the reading stabilizes.

11.3.1 Insert the LBOD probe into the BOD bottle of the prepared samples for BOD or cBOD determination. Insure that there are no air bubbles that may have collected around the probe or sensor. Turn on the stir paddle and read sample. The display will show "Stabilizing" and a progress bar as the probe stabilizes in the sample. The display will show the lock icon when the reading stabilizes.

12.0 Procedure for the Preparation and Determination of BOD₅ and cBOD₅ Samples

12.1 Follow the instrument manufacturer's instructions for instrument setup for the LBOD probe (Hach Document DOC022.53.80025).

12.2 The BOD test is a 5-day test. Follow all steps carefully to make sure that the test does not have to be repeated.

12.3 The dilution water for this test must be fully air-saturated and not have an oxygen demand or any toxins. When incubated for 5 days at 20 °C, the dissolved oxygen concentration in the dilution water must not change by more than 0.2 mg/L.

12.4 Distilled Water Preparation

12.4.1 The dilution water must be prepared very carefully to make sure that no source of oxygen demand or toxins are added. The water that is used to prepare the dilution water must be of very high quality. The water must not have any organic compounds or any toxic compounds such as chlorine, copper, and mercury at a concentration level that would interfere would the BOD seed and inhibit microbiological growth of organisms.

- 12.4.2 For best results, use an alkaline permanganate distillation for preparing dilution water. Resin in ionization cartridges will occasionally release organic materials that have an oxygen demand.
- 12.4.3 Store the distilled water in clean jugs at a temperature of 20 °C. Fill the containers to about $\frac{3}{4}$ full and shake the jugs to saturate the water with air. Alternatively, saturate the water with air as described in Section 7.2.1. A small aquarium pump or air compressor can be used to saturate the water with air. Insure that the air is filtered and that the air filter does not grow bacteria.

12.5 *Dilution Water Preparation*

- 12.5.1 Using the distilled water prepared above, select a BOD nutrient buffer pillow from the BOD nutrient buffer pillows table.
- 12.5.2 Add the contents of the BOD Nutrient Buffer Pillow to the distilled water in a jug with ample headspace. Cap the jug and shake vigorously for one minute to dissolve the nutrients and to saturate the water with air.
- 12.5.3 Alternatively, prepare the dilution water by adding 1 mL each of the following solutions per liter of distilled water prepared in Section 12.4.
- Buffer Solution, APHA, for BOD, pH 7.2, Phosphate type, (Hach Catalog Number 43149)
- Calcium Chloride Solution, APHA, for BOD (Hach Catalog Number 42849)
- Ferric Chloride Solution, APHA, for BOD (Hach Catalog Number (42953)
- Magnesium Sulfate Solution, APHA, (Hach Catalog Number (43094)
- 12.5.4 Cap the jug and shake vigorously for one minute to dissolve the nutrients and to saturate the water with air.

12.6 *Seed Preparation*

- 12.6.1 Use raw sewage or other reliable sources for the bacterial seed that will yield 198 ± 30 mg/L BOD with the GGA check sample in Section 9.6. Potential seed sources include wastewater influent, primary effluent, soil, and domestic sewage.
- 12.6.2 Allow raw sewage to stand undisturbed at 20 °C for 24 to 36 hours before use.
- 12.6.3 When seeding samples with raw sewage, always pipette from the upper portion of the sewage.

12.7 *Sample Size Selection Guide*

- 12.7.1 Make an estimate of the sample volumes that are necessary for the test. At least 2.0 mg/L of DO should be consumed during the test and at least 1.0 mg/L of un-depleted DO should remain in the bottle.
- 12.7.2 Samples such as raw sewage will have a high BOD. Small sample volumes must be used because large samples will deplete all of the oxygen in the sample. Samples with a low BOD must use larger sample volumes to insure that adequate oxygen is depleted to give accurate results.

- 12.7.3 Refer to the Minimum Sample Volume Table to select the minimum sample volume. For example, if a sewage sample is estimated to contain 300 mg/L BOD, the minimum sample volume is 2 mL. For sewage effluent with an estimated BOD of 40 mg/L, the minimum sample volume is 15 mL.
- 12.7.4 Refer to the Maximum Sample Volume Table to select the maximum sample volume. At 1000 in elevation, with an estimated BOD of 300 mg/L, the largest sample volume is 8 mL. For a BOD of 40 mg/L, the maximum volume of sample is 60 mL.

12.8 *Sample Matrix Pretreatment*

12.8.1 Determine the pH of each sample prior to BOD sample preparation. For samples that of have pH of less than 6.0 or greater than 8.5, adjust the pH accordingly with a solution of sulfuric acid (H₂SO₄) or sodium hydroxide (NaOH). Strength of pH adjustment solution should be at a concentration that does not dilute the sample by greater than 0.5 percent.

12.8.2 For sample matrices that contain residual chlorine, de-chlorinate with a solution of (Na₂S₂O₃).

12.8.2.1 Measure 100 mL of sample into a 250 mL Erlenmeyer flask. Using a 10-mL serological pipette and pipette filler, add 10 mL of 0.020 N Sulfuric Acid Standard Solution and 10 mL of Potassium Iodide Solution, 100-g/L, to the flask.

12.8.2.2 Add three full droppers of Starch Indicator Solution and swirl to mix.

12.8.2.3 Fill a 25-mL burette with 0.025 N Sodium Thiosulfate Standard Solution and titrate the sample from dark blue to colorless.

12.8.2.4 Calculate the amount of 0.025 N Sodium Thiosulfate Solution to add to the sample:

$$\text{mL } 0.025 \text{ N Sodium Thiosulfate required} = \text{mL titrant used} \times \text{volume of remaining sample divided by } 100$$

12.8.2.5 Add the required amount of 0.025 N Sodium Sulfate Standard Solution to the sample. Mix thoroughly and wait 10 to 20 minutes before performing the BOD test.

12.9 *Sample Preparation*

12.9.1 Select the sample volume as described in Section 12.7. Select a minimum of three different volumes for each sample.

12.9.1.1 If the minimum sample volume is 3 mL or more, determine the DO concentration in the undiluted sample; this determination can be omitted when analyzing sewage and settled effluents known to have a dissolved oxygen content near 0 mg/L.

12.9.2 Stir the sample gently with a pipette. Use the pipette to add the determined sample volumes to the BOD bottles.

12.9.3 Add the appropriate seed to the individual BOD bottles as described in Section 12.6.

- 12.9.3.1 Separately, with each batch of BOD samples, prepare a seed sample with dilution water. Measure the BOD of the seed for subtraction from the sample BOD.
- 12.9.3.2 A seed that has a BOD of 200 mg/L (a typical range for domestic sewage) will typically deplete at least 0.6 mg/L DO when added at a rate of 3 mL/L of dilution water.
- 12.9.4 If the test is for cBOD, add two portions of Nitrification Inhibitor (approximately 0.16 g) to each bottle. The oxidation of nitrogen-based compounds will be prevented.
- 12.9.5 Fill each bottle to just below the lip with dilution water.
 - 12.9.5.1 Allow the dilution water to flow down the sides of the bottle to prevent air bottles from becoming entrapped in the bottle.
- 12.9.6 Fill an additional BOD bottle with only dilution water. This will be the dilution water blank.
- 12.9.7 Stopper the bottles carefully to prevent air bubbles from becoming entrapped.
 - 12.9.7.1 Tightly twist the stopper and invert the bottles several times to mix.
- 12.10 *Sample Analysis*
 - 12.10.1 Measure the initial dissolved oxygen concentration in each bottle with the LBOD probe.
 - 12.10.2 After the initial DO measurement, stopper the bottles carefully to prevent air bubbles from becoming entrapped.
 - 12.10.2.1 Add dilution water to the lip of each BOD bottle to make a water seal.
 - 12.10.3 Place a plastic cap over the lip of each bottle and incubate at 20 ± 1 °C for five days.
 - 12.10.4 After 5 days, measure the remaining dissolved oxygen concentration in each bottle with the LBOD probe.
 - 12.10.4.1 At least 1.0 mg/L DO should have remained in each bottle.
 - 12.10.4.2 Discard results of samples where the DO is depleted below 1. mg/L.

13.0 BOD and cBOD Calculations

13.1 *When Dilution Water Not Seed (generally influent and primary treated influent to treatment)*

$$\text{BOD}_5 \text{ or cBOD}_5, \text{ mg/L} = \frac{D_1 - D_2}{P}$$

where:

BOD₅ or cBOD₅ = BOD value from the 5-day test

D₁ = DO of diluted sample immediately after preparation, in mg/L

D₂ = DO of diluted sample after 5 day incubation at 20 °C, in mg/L

P = Decimal volumetric fraction of sample used

f = ratio of seed in diluted sample to seed in seed control (% seed in diluted sample/%seed in seed control) or, if seed material is added directly to sample or to seed-control bottles:

f = ((volume of seed in diluted sample/volume of seed in seed control)

13.2 When Dilution Water Requires Seed

$$\text{BOD}_5 \text{ or cBOD}_5, \text{ mg/L} = \frac{(D_1 - D_2) - (B_1 - B_2)f}{P}$$

where as defined above plus:

B_1 = DO of seed control before incubation, in mg/L

B_2 = DO of seed control after incubation, in mg/L

13.3 Averaged Results

13.3.1 Averaged results from different dilutions are acceptable if more than one sample dilution meets all of the following criteria:

13.3.1.1 The remaining un-depleted DO is at least 1 mg/L.

13.3.1.2 The final DO value is at least 2 mg/L lower than the initial prepared sample DO

13.3.1.3 There is no evidence of toxicity at higher sample concentrations

14.0 Method Performance for Dissolved Oxygen in Reference Water and GGA Recovery

Acceptance Criterion	Section	Limit
Initial DO Accuracy in Reagent Water	9.2.1	95% to 105%
Initial Precision in Reagent Water	9.2.1	2.1%
Average GGA Recovery	15.7	198 ± 30.5 mg/L BOD

15.0 Pollution Prevention

15.1 There are no standards or reagents used in this method when properly disposed of, pose any threat to the environment.

14.0 Waste Management

14.1 It is the laboratory's responsibility to comply with all federal, state, and local regulations governing waste management, particularly the hazardous waste identification rules and land disposal restrictions, and to protect air, water, and land by minimizing and control all releases from fume hoods and bench operations. Compliance with all sewage discharge permits and regulations is also required.

14.2 For further information on waste management, consult "The Waste Management manual for Laboratory Personnel", and Less is Better: Laboratory Chemical Management for Waste Reduction", both available from the American Society's Department of Government relations and Science Policy, 1155 16th Street N.W., Washington, D.C. 20036.

15.0 References

15.1 Handbook of Analytical Quality Control in Water and Wastewater Laboratories," USEPA, EMSL-CI, Cincinnati, OH 45268, EPA-600-4-79-019, March 1979.

15.2 Hitchmen, M.L. (1978) *Chemical analysis*. Vol. 49. *Measurement of Dissolved Oxygen*. Wiley and sons, New York.

15.3 Title 40, Code of Federal Regulations (40 CFR), Part 136.

- 15.4 Protocol for EPA Approval of New Methods for Organic and Inorganic Analytes in Wastewater and Drinking Water (EPA-821-B-98-003, March 1999).
- 15.5 "OSHA Safety and Health Standards, General Industry," (29 CFR 1910), Occupational Safety and Health Administration, OSHA 2206 (Revised, January 1976)
- 15.6 "Safety in Academic Chemistry Laboratories," American Chemical Society, Committee on Chemical Safety, 3rd Edition, 1979.
- 15.7 "Standard Methods for the Examination of Water and Wastewater", 20th Edition; American Public health Association: 1015 Fifteenth Street, NW, Washington, D.C. 2005, 1998; Method 5210B.

16.0 Tables

16.1 Nutrient Buffer Preparation Options

Table 1 - BOD Nutrient Buffer Pillows

Volume of Dilution Water to Prepare	Hach BOD Nutrient Pillow Catalog Number
300 mL) add pillow to each BOD Bottle	1416066
3 liters	1486166
4 liters	2436466
6 liters	1486266
19 liters	1486398

Table 2 – Conventional Dilution Water Preparation Reagents

Hach Buffer Reagents	Hach Buffer Reagent Catalog Number
Buffer Solution, APHA, for BOD, pH 7.2, phosphate type, 500 mL	43149
Calcium Solution, Alpha, for BOD, 500 mL	42849
Ferric Chloride Solution, APHA, 1000 mL	42953
Magnesium Sulfate Solution, APHA, for BOD, 500 mL	43049

Note: The Phosphate Buffer Solution should be refrigerated to decrease the rate of biological growth.

16.2 Sample Volume Selection Guides

Table 3 - Minimum Sample Volume Selection Guide

Sample Type	Estimated BOD mg/L	Minimum Sample Volume (mL)
Strong Waste	600	1
Raw and Settled Sewage	300	2
	200	3
	150	4
	100	6
	75	8
	60	10
	50	12
Oxidized Effluents	40	15
	30	20
	20	30
	10	60
	6	100
Polluted River Water	4	200
	2	300

Table 4 – Maximum Sample Volume Selection Guide

BOD at Sea Level (mg/L)	BOD at 1000 ft Elevation (mg/L)	BOD at 5000 feet Elevation (mg/L)	Maximum Sample Volume (mL)
615	595	508	4
492	476	406	5
410	397	339	6
304	294	251	8
246	238	203	10
205	198	169	12
164	158	135	15
123	119	101	20
82	79	68	30
41	40	34	60
25	24	21	100
12	12	10	200
8	8	7	300

Note: Samples with higher concentrations of BOD should be pre-diluted.

16.3 Performance Criteria

Table 5 - Initial Precision and Recovery Method Performance

IPR Range	IPR DO Conc. (mg/L)	97.5% Lower Limit of Recovery (%)	97.5% Upper Limit of Recovery (%)	95% Upper Limit of Precision (%)
Low	1.72 – 1.74	95.0	104.0	2.07
High	7.22 – 9.23	95.8	104.8	1.26

Table 6 - Calibration Verification Performance

CV DO Concentration	Average % Recovery	% Standard Deviation	% Relative Standard Deviation
7.22 mg/L – 9.23 mg/L	100.1	2.5	2.5

17.0 Glossary of Definitions and Purposes

The definitions and purposes are specified to this method but have been conformed to common usage as much as possible.

17.1 Units of Weight and Measure and their Abbreviations

17.1.1 Symbols
°C degrees Celsius

17.1.2 Alphabetical characters
mg/L milligram per liter

17.2 Definitions, acronyms, and abbreviations

17.2.1 LDO® - Luminescence dissolved oxygen

17.2.2 BOD₅: Biological oxygen demand, 5-day test

17.2.3 cBOD₅: Carboneous biological oxygen demand, 5-day test

17.2.4 DO: Dissolved oxygen

17.2.5 CV: Calibration verification

17.2.6 IPR: Initial precision and recovery

17.2.7 OPR: On-going precision and recovery